



# Certificate of Compliance

**Certificate:** 80036174

**Master Contract:** 301072

**Project:** 80036174

**Date Issued:** 2020-04-29

**Issued To:** Illumina, Inc  
5200 Illumina Way  
San Diego, California, 92122  
United States

**Attention:** Francis Hwang

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*

**Issued by:** *Christoph Choudhuri*  
Christoph Choudhuri



## **PRODUCTS**

CLASS – 8721 06 - ELECTRICAL LABORATORY EQUIPMENT

CLASS – 8721 86 - ELECTRICAL LABORATORY EQUIPMENT Certified to US Standards

**Multichannel Pipetting Instrument**, Model INFINIUM LiHa, INFINIUM RoMa,  
rated AC 100 – 120 V / 220 – 240 V, 50/60 Hz, 1200 VA, all models are cord connected (detachable cord set),  
Equipment Class I (PE connected), Overvoltage category II, Pollution degree 2, max. 2000 m operating altitude,  
+15 °C to +32 °C ambient operating temperature.



**Certificate:** 80036174  
**Project:** 80036174

**Master Contract:** 301072  
**Date Issued:** 2020-04-29

### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018	- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
CSA C22.2 No. 61010-2-010:19	- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of materials
CSA C22.2 No. 61010-2-081:19	- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-081: Particular Requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes
UL 61010-1, 3rd Edition (2012), AMD1:2018	- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
UL 61010-2-010:2019 (Fourth Edition)	- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of materials
UL 61010-2-081:2019 (Fourth Edition)	- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-081: Particular Requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

### **Conditions of Acceptability:**

1. The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
2. Equipment is only to be installed by manufacturer trained personnel.
3. Equipment is not to be used with flammable liquids.
4. The accuracy of measured, analyzed or prepared quantities has not been investigated.
5. External waterbath (heating and cooling) intended to be used with the equipment is not part of this certification.



## *Supplement to Certificate of Compliance*

**Certificate:** 80036174

**Master Contract:** 301072

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80036174	2020-04-29	Initial cCSAus Certification for Multichannel Pipetting Instrument, Models Infinium ROMA and Infinium LIHA based on CB Acceptance from Eurofins.