

Certificate



Certificate no.

CU 72143120 02

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte. Ltd.
North Tech Lobby 3 #02-13/18
29 Woodlands Industrial Park E1
Singapore 757716
Singapore

Test report no.: USA-DR 31480867 005

Client Reference: Carol Escano

Tested to:

UL 61010-1:2004 R10.08
IEC 61010-2-010:2003
IEC 61010-2-081:2001+A1
CAN/CSA-C22.2 NO. 61010-1-04+GI1 (R2009)
CAN/CSA-C22.2 NO. 61010-2-010-04 (R2009)
CAN/CSA-C22.2 No. 61010-2-081-04 (R2009)

Certified Product: Library Prep System

License Fee - Units

Addition:

Manufacturing Plant: see above (K754252)

Rated Ambient Temperature: 25°C

Appendix: 1, 1-4

Licensed Test mark:



Date of Issue

(day/mo/yr)

23/03/2016