



Ref. Certif. No.

DE 2-038501-A1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Genetic Sequencer
Name and address of the applicant	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the manufacturer	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the factory	Illumina Singapore Pte. Ltd. #02-13/18 29 Woodlands Industrial Park E1. North Tech Lobby Singapore 757716, Singapore
Ratings and principal characteristics	AC 100-240V, 50/60Hz; 600W; Class I; Laser Class 1
Trademark (if any)	illumina
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NextSeq 500, NextSeq 550
Additional information (if necessary may also be reported on page 2)	Reissue of CB Certificate DE 2-038501 dated 30.01.2023 due to modification (update of critical component list). See associated test report for details.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 IEC 61010-2-010:2019 IEC 61010-2-081:2019 IEC 60825-1:2014 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31393108 044

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2023-09-15

Signature:

James Howell